





AIR RING ASPIRATION





ASPIR 2 QC



Code: AS-100-QC

Description: Suction Unit 1 Hp complete with electronic control board and rubber feet.

| |  |  |  |  |
|------------|--|--|--|--|
| Model | Dental chairs | Air Capacity | Motor | Volt. Freq. |
| ASPIR 2 QC | 2 | <i>m3/h</i> <i>l/min</i> <i>cfm</i> 2.280 80 | <i>HP</i> <i>KW</i> 1 0,75 | V Hz 230 50/60 |

| |  |  |  |  |
|------------|---|---|---|---|
| Model | Noise Level | Max oper. head | Size | Net weight |
| ASPIR 2 QC | <i>dB(A)</i> 62 | <i>mbar</i> 140 | <i>mm</i> 330x290x340 | <i>Kg</i> <i>lb</i> 18 40 |